

**GUAM MEMORIAL HOSPITAL AUTHORITY  
850 GOV. CARLOS G. CAMACHO ROAD  
TAMUNING, GUAM 96913  
PHONE: 671-647-2165  
EMAIL: [materials.mgmt@gmha.org](mailto:materials.mgmt@gmha.org)**

**TRANSMITTAL SHEET**

**TO:** Prospective Bidders  
**FROM:** Hospital Materials Management Administrator  
**DATE:** September 11, 2024  
**SUBJECT:** **IFB-020-2024- Purchase of Fluoroscopy X-Ray System for the GMHA Radiology Department. Amendment No. 3**  
**PAGES:** 2 including cover sheet

**NOTES:**

An acknowledgement via a return email would be appreciated as soon as possible.

**DATE / VENDOR ACKNOWLEDGEMENT**

**FACSIMILE CONFIDENTIALITY NOTICE**

These documents accompanying this transmission contain confidential information that is legally privileged. This information is intended only for the use of the individual or entity named above. The authorized recipient of this information is prohibited from disclosing this information to any other party unless required to do so by law or regulation and is required to destroy the information after its stated need has been fulfilled.

If you are not the intended recipient, you are hereby notified that any disclosure, copying, distribution, or action taken in reliance on the contents of these documents is strictly prohibited. If you have received this information in error, please notify the sender immediately and arrange for the return or destruction of these documents.



# GUAM MEMORIAL HOSPITAL AUTHORITY

ATURIDÁT ESPETÁT MIMURIÁT GUÅHÅN

850 Governor Carlos Camacho Road, Tamuning, Guam 96913  
Operator: (671) 647-2330 or 2552 | Fax: (671) 649-5508



## AMENDMENT #3 GMHA IFB 020-2024 Purchase of Fluoroscopy X-Ray System

This amendment is in response to clarification and request from Medpharm.

**Clarification:** Continuous Fluoroscopy time: Minimum 10 minutes. This requirement is apparently a mistake, please check with end-user the maximum Continuous Fluoroscopy Exposure.

**Response:** The specification of a minimum 10-minute continuous fluoroscopy time is indeed an error. The correct requirement is the maximum continuous fluoroscopy time, which is typically 5 minutes before the system automatically pauses for safety.

**Medpharm Clarification:** While we strive our best with inventory of recommended spare parts, the required backup unit to be provided on the 4<sup>th</sup> day downtime will be extremely difficult and costly.


**Response:** GMHA may forego a loaner or backup as long as the contractor provides the recommended parts by the OEM for spare parts list.

**Medpharm Clarification:** Providing Level II Biomedical Training – Manufacturer's in most cases restricts service training to authorized distributors only. If and when 3<sup>rd</sup> party or customers are to be accommodated, limited to Level I only.

**Response:** GMHA will accept Level I service training if and when available as Level II Biomedical Training is restricted to authorized distributors only. Total bid amount must be inclusive of Level I training.

Please acknowledge receipt of this amendment by signing and sending back to Materials Management by Fax at 649-3640 or email to [materials.mgmt@gmha.org](mailto:materials.mgmt@gmha.org).

If you have any questions, please feel free to address your letter to Lillian Perez-Posadas MN, RN, and fax to the Materials Management Office at 649-3640 or email to the [materials.mgmt@gmha.org](mailto:materials.mgmt@gmha.org).

Sincerely:  


DOLORES PANGELINAN  
Hospital Materials Management Administrator

ACKNOWLEDGMENT:

\_\_\_\_\_  
PRINT NAME

\_\_\_\_\_  
SIGNATURE

\_\_\_\_\_  
DATE